



Part No. 74000 Pre-Surgical Scope Warmer  
**MATERIAL SAFETY DATA SHEET**

**Product Name: Sodium Acetate**

EPA Reg No. N/A

**1. PRODUCT IDENTIFICATION**

Product Name.....Sodium Acetate

UN/MA#.....N/A

DOT Hazard Class.....N/A

**2. INGREDIENTS & RECOMMENDED OCCUPATIONAL EXPOSURE LIMITS**

Sodium Acetate is non-hazardous

**3. PHYSICAL DATA**

Specific Gravity (Hg0=1).....1.53

Boiling Point.....N/A

Vapor Pressure.....N/A

Vapor Density.....N/A

pH.....8-10.5

Solubility in Water.....Partially soluble in water

Evaporation Rate.....N/A

Appearance and Odor.....White to Grey angular granule

**4. FIRE AND EXPLOSION HAZARD DATA**

Flash Point.....>100°C

Flammable Limits.....NDA

Fire Extinguishing Media.....Use extinguishing media appropriate for surrounding fire



## Part No. 74000 Pre-Surgical Scope Warmer

Special Fire Fighting Procedures.....N/A

Unusual Fire and Explosion Hazards.....None

### 5. REACTIVITY DATA

Stability.....Stable

Incompatibility.....Strong oxidizing agents

Hazardous Decomposition or

By-Products.....Carbon monoxide

Conditions to Avoid .....N/A

Hazardous Polymerization.....Will not occur

### 6. SPILL OR LEAK PROCEDURES

Steps to be taken in case material is released: In case of release to the environment, report spills to the National Response Center 1-800-424-8802

Suggested Local Action: Dike spill and pump to clean container. Flush area with water

Waste Disposal Method: (EPA Waste Identification #: N/A – Consult Federal, State & Local Regulations) In accordance with all Federal, State and Local regulations.

For Hazardous Waste Regulations: call 1-800-424-9346 – The RCRA Hotline.

### 7. HEALTH HAZARD DATA

Inhalation: Possible with ensuing irritation.

Skin Contact: Possible with ensuing irritation.

Eye Contact: Possible with ensuing irritation.

Ingestion: Possible with ensuing irritation.

Signs & Symptoms of Exposure: Irritation of skin, eyes and respiratory tract.

ACUTE: Irritant



## Part No. 74000 Pre-Surgical Scope Warmer

CHRONIC: N/A

Medical conditions generally aggravated by exposure; Irritation of respiratory tract.

Chemical listed as Carcinogen or Potential Carcinogen: No

### 8. EMERGENCY AND FIRST AID PROCEDURE

Skin contact: Wash thoroughly with soap and water.

Eye contact: Flush immediately with large amounts of water, lifting the lower and upper lids occasionally.

Ingestion: Give 1-2 large glasses of water or milk. Induce vomiting.

Immediately seek medical aid. Never give liquids to an unconscious person.

### 9. SPECIAL PROTECTION INFORMATION

Skin: Clean, body-covering clothing should be worn to prevent irritation in situation where direct contact with product may occur.

Eyes: Employees should be required to wear chemical safety splash goggles in situations where direct contact with the product may result in eye injury.

Other Protective Equipment: Emergency eye wash stations and deluge safety showers should be available in the work areas.

### 10. SPECIAL PRECAUTIONS

Precaution to be taken in handling and storage: Store in a cool, dry location. DO NOT store near food or feed. Keep out of reach of children and pets.

THIS INFORMATION IS TAKEN FROM SOURCES OR BASED UPON DATE BELIEVED TO BE RELIABLE:

**Product Name.....Sodium Acetate**

#### Materials

#### CAS# OSHA-PEL ACGLH-TLV

Sodium Acetate Anhydrous Water-Insoluble 127-09-3      N/A N/A



MEDIFLEX Division of Flexbar Machine Corporation  
250 Gibbs Road, Islandia, NY 11749-2697 • USA

SURGICAL DEVICES SINCE 1969

Phone: 631-582-8440 • Fax 631-582-8487

Toll Free: 1-800-879-7575

[www.mediflex.com](http://www.mediflex.com) • e-mail: [sales@mediflex.com](mailto:sales@mediflex.com)



## Part No. 74000 Pre-Surgical Scope Warmer

# SGS

**Test Report No.** F690501/LF-CT6GP06-03371

**Date:** February 24,2006

Sample No. : GP06-033371.001  
Sample Description : Stainless Steel Strip 304  
Style/Item No. : N/A  
Comments : Material is stainless steel strip 304

### Heavy Metals

Test Items	Unit	Test Method	MDL	Results
Cadmium(Cd)	mg/kg	US EPA 3050B(1996).US EPA 6010B(1996)	0.5	ND
Lead(Pb)	mg/kg	US EPA 3050B(1996).US EPA 6010B(1996)	5	ND
Mercury(Hg)	mg/kg	US EPA 3052(1996).US EPA 6010B(1996)	2	ND
Hexavalent Chromium (Cr Vi)	mg/kg	US EPA 3060A(1996).US EPA 7196A(1992)	1	ND
Arsenic(As)	mg/kg	US EPA 3050B(1996).ICP AES	10	ND
Barium(Ba)	mg/kg	US EPA 3050B(1996).ICP AES	0.5	ND
Antimony(Sb)	mg/kg	US EPA 3050B(1996).ICP AES	10	ND
Selenium(Se)	mg/kg	US EPA 3050B(1996).ICP AES	10	ND

NOTE: (1) ND = Not Detected. (<MDL)  
(2) ppm =mg/kg  
(3) MDL = Method Detection Limit  
(4) Negative = Undetectable /Positive = Detectable



MEDIFLEX Division of Flexbar Machine Corporation  
250 Gibbs Road, Islandia, NY 11749-2697 • USA

SURGICAL DEVICES SINCE 1969

Phone: 631-582-8440 • Fax 631-582-8487  
Toll Free: 1-800-879-7575  
[www.mediflex.com](http://www.mediflex.com) • e-mail: [sales@mediflex.com](mailto:sales@mediflex.com)



Part No. 74000 Pre-Surgical Scope Warmer

ITEMS	UNIT	TEST RESULTS	TEST METHOD
Lead (Pb)	mg/kg	18	ASTM D 4004 93
Cadmium(Cd)		Less than 5	
DBP		Not detected (detection limit 50)	
DIDP		Not detected (detection limit 50)	
BBP		Not detected (detection limit 50)	
DEHA		Not detected (detection limit 50)	
DEHP		Not detected (detection limit 50)	
DNOP		Not detected (detection limit 50)	
DINP		Not detected (detection limit 50)	

\*Pharote Free P.V.C.



MEDIFLEX Division of Flexbar Machine Corporation  
 250 Gibbs Road, Islandia, NY 11749-2697 • USA

SURGICAL DEVICES SINCE 1969

Phone: 631-582-8440 • Fax 631-582-8487  
 Toll Free: 1-800-879-7575

[www.mediflex.com](http://www.mediflex.com) • e-mail: [sales@mediflex.com](mailto:sales@mediflex.com)